



Test Monitoring Center

@ Carnegie Mellon University
6555 Penn Avenue, Pittsburgh, PA 15206, USA

<http://astmtmc.cmu.edu>
412-365-1000

MEMORANDUM: 15-014
DATE: May 6, 2015
TO: Gil Reinhard, Chairman, CBT Surveillance Panel
FROM: Michael T. Kasimirsky *Michael T. Kasimirsky*
SUBJECT: HTCBT Testing from October 1, 2014 through March 31, 2015

A total of 254 HTCBT tests were reported to the Test Monitoring Center during the report period from October 1, 2014 through March 31, 2015.

Please find attached a summary of testing activity this period.

MTK/mtk/astm1015.doc/mem15-014.mtk.doc

cc: F. M. Farber

J. A. Clark

CBT Surveillance Panel


<ftp://ftp.astmtmc.cmu.edu/docs/bench/htcvt/semiannualreports/htcvt-04-2015.pdf>

Distribution: email

HTCBT (D 6594)

Reporting Data	
Number of Labs	6

NUMBER OF TESTS REPORTED
BY LAB AND REPORT PERIOD



14:16:50 05MAY2015


HTCBT (D 6594)

Test Distribution by Validity

		Number of Tests
Accepted for calibration	AC	230
Rejected	OC	9
Invalidated	LC	1
Aborted	XC	1
Acceptable Shakedown Run	AG	5
Unacceptable Shakedown Run	OG	6
Invalid Shakedown Run	LG	2
Aborted Shakedown Run	XG	0
Total		254

HTCBT (D 6594)

CALIBRATION ATTEMPT SUMMARY




Resolution

Report Period

14:16:50 06MAY2015

HTCBT (D 6594)

OPERATIONALLY VALID TESTS
MEETING ACCEPTANCE CRITERIA



14:16:50 05MAY2015

HTCBT (D 6594)

CAUSES FOR LOST TESTS

Summary of Reasons for Failed Tests	No. of Tests
Copper, severe	0
Copper, mild	2
Lead, severe	3
Lead, mild	1
Copper & Lead, severe	3

HTCBT (D 6594)

CAUSES FOR LOST TESTS (CONTINUED)

Summary of Reasons for Invalid Tests	No. of Tests
Sample Contamination	1

HTCBT (D 6594)

CAUSES FOR LOST TESTS (CONTINUED)

Summary of Reasons for Aborted Tests	No. of Tests
Airflow Problem	1

HTCBT (D 6594)

Average Δ /s By Laboratory			
Lab	n	CUC	PBC
A	58	1.274	0.757
B	70	0.887	-0.285
G	34	1.071	1.238
I	67	0.193	0.198
P	10	0.915	-0.691
V	0	-	-
Industry	239	0.814	0.303

Individual test results can be found on the TMC Web Page at the following link:

<ftp://ftp.astmtmc.cmu.edu/refdata/bench/htcbt/data/>


HTCBT (D 6594)

HIGH TEMP CBT INDUSTRY OPERATIONALLY VALID DATA



COPPER CHANGE (ppm)

CUSUM Severity Analysis



06MAY15: 14:17




HTCBT (D 6594)

HIGH TEMP CBT INDUSTRY OPERATIONALLY VALID DATA



LEAD CHANGE (ppm)

CUSUM Severity Analysis




06MAY15: 14:17



HTCBT (D 6594)

TEST PRECISION

POOLED STANDARD DEVIATION
BY SIX-MONTH ASTM REPORT PERIOD



14:16:50 05MAY2015

HTCBT (D 6594)

SUMMARY OF SEVERITY & PRECISION

Severity

Over the course of this report period, copper severity, as measured by cusum plotting, was severe.

Over the course of this report period, lead severity, as measured by cusum plotting, was severe.

Precision

Pooled s for this period is 0.22 for copper and 7.38 for lead.

Over the course of this report period, Precision, as measured by pooled standard deviation, is unchanged from last period for copper and is slightly worse than the previous period for lead, but is still within historical levels.

HTCBT (D 6594)

INFORMATION LETTERS

No HTCBT Information Letters were issued this period.

HTCBT (D 6594)

STATUS OF REFERENCE OIL SUPPLY

Reference Oil	Samples @ Labs	@ TMC	
		Samples (4 oz)	Gallons
44-1	0	0	0.0
44-2	0	0	0.0
44-3	71	316	9.9
1005-1	0	0	0.0
1005-3	116	86	2.7
Total	187	402	12.6

A reblend of 1005-3, reference oil 1005-5, is available at the TMC, but has not been introduced into HTCBT testing.

The TMC is working on procuring a reblend of reference oil 44-3; it should be available in July 2015.