



Test Monitoring Center

6555 Penn Avenue
Pittsburgh, PA 15206-4489
(412) 365-1000

MEMORANDUM: 01-025
DATE: March 19, 2001
TO: TEOST Participants
FROM: Tom Schofield
SUBJECT: MHT-4 TEOST Reporting Package Clarification (Version 20010208)

It has been brought to my attention that a field in the latest MHT-4 TEOST reporting package may not be represented properly and this has led to some confusion by users of the package. The data dictionary in report package version 20010208 indicates that the Net Sample Weight of Injected Catalyzed Sample, g (SAMINJWT) is to be reported to four decimal places, while Section 11.8.6 of the procedure (Draft 17, 00.08.11 is the latest I have) says the syringe only needs to be weighed to the nearest 0.01 g. The report forms are not intended to redefine the procedure, and the syringe weight only needs to be reported to a minimum of two decimal places. If you wish to report the injected sample weight more precisely, you may do so.

Please direct all inquiries to the TMC.

TMS/tms

c: Ted Selby, Savant, Inc.
J. Zalar, TMC
<ftp://tmc.astm.cmri.cmu.edu/docs/bench/mteos/memos/memo01-025.tms.pdf>

Distributed by e-mail