



Test Monitoring Center

Carnegie Mellon University
6555 Penn Avenue, Pittsburgh, PA 15206, USA

<http://astmtmc.cmu.edu>
412-365-1000

ISB Information Letter 12-2
Sequence No. 6
August 30, 2012

ASTM consensus has not been obtained on this information letter. An appropriate ASTM ballot will be issued in order to achieve such consensus.

TO: Cummins Mailing List
SUBJECT: ISB Calibration Period Change

On August 28, 2012, via teleconference, the Cummins Test Surveillance panel unanimously approved reducing the calibration period for ISB test stands from 18 months to 12 months. This change is effective with any calibration period established on or after August 29th, 2012. Section 9.3 of D7484 has been revised and is attached.

Shawn Whitacre
Director – Chemical Technology
Cummins Inc.

Frank M. Farber
Director
ASTM Test Monitoring Center

Attachment
c: ftp://ftp.astmtmc.cmu.edu/docs/diesel/cummins/procedure_and_ils/isb/il12-2.pdf

Distribution: Email

(Revises D 7484-11, as modified by Information Letter 12-1)

9.3 Stand Calibration Period:

9.3.1 A test stand calibration test is carried out 12 months or 12 operationally valid non-reference oil tests, whichever comes first, following the completion of the last successful calibration test.

Delete current 9.3.2

9.3.2 *Last Start Date*—A non-reference oil test can be completed provided the warm-up is started prior to the expiration of the calibration period.