ISM Update

Shawn D. Whitacre June 29, 2010

HDEOCP Kansas City, MO



Attachment 5: Page 1 of 5



Outline

- Oil Filters
- Crossheads
- Injector Adjusting Screws





Oil Filters

- New batch of ISM oil filters introduced in 2009
- Were never run on reference test, but concerns expressed based on results of some candidate runs
- Cummins Filtration produced new batch in Q1 2010 with retargeted media
- "2010 batch" used in series of reference tests
- Results very much in line with historical data
- "2010 filters" approved; 2009 filters sequestered and will not be used. 5+ year supply of 2010 filters on-hand.





Crossheads

- New batch (D) of crossheads introduced for recent series of reference tests
- Results in-line with historical data, but trending slightly mild
- SP will monitor trend and consider CF adjustment as necessary
 - Motion to adjust CF based on existing data failed





Injector Adjusting Screw

- IAS weight loss correction factor (+19.1 mg) June 28, 2007
- CF still statistically appropriate but impact on merit system has been called in question
 - Full merits (16 mg min) are no longer in play
- SP deliberating a variety of proposals but none that are ready to bring to HDEOCP