

Liner Measurement Round Robin Winter 2003

LINER: TMC 480634

Reference Test Measured Value: 27.7 microns

| <u>Lab</u> | <u>PDI</u> | <u>Trial</u> | | | | <u>Average**</u> |
|------------|------------|--------------|----------|----------|----------|------------------|
| | | <u>1</u> | <u>2</u> | <u>3</u> | <u>4</u> | |
| A | 1 | 28.36 | 28.40 | 28.23 | 28.33 | 28.33 |
| A | 2 | 28.78 | 28.88 | 28.72 | 28.85 | 28.81 |
| A | 3 | 28.49 | 28.38 | 28.52 | 28.51 | 28.47 |
| A | 4 | 28.54 | 28.57 | 28.55 | 28.55 | 28.55 |
| B | 1 | 28.73 | 28.77 | 28.46 | 28.59 | 28.64 |
| D | 1 | 28.44 | 28.50 | 28.41 | 28.52 | 28.47 |
| F | 1 | 28.45 | 28.49 | 28.51 | 28.66 | 28.53 |
| G | 1 | 29.14 | 29.16 | 29.15 | 29.22 | 29.17 |

| | |
|--------------------|-------|
| Target | 28.62 |
| Standard Deviation | 0.24 |
| Minimum | 27.90 |
| Maximum | 29.34 |

**Average values are for information only. Calibration is determined by comparing individual measurements to the Minimum and Maximum values.