

# IVB Metrology Sub-Group | MINUTES

REVISION DATE: 7/2/2018 2:50:00 PM

<b>Relevant Test:</b>	Sequence IVB
<b>Note Taker:</b>	Chris Mileti
<b>Meeting Date:</b>	06-27-2018
<b>Comments:</b>	This was the "kick-off" conference call for the Sequence IVB Metrology Sub-Group.

## 1. REVIEW OF AGENDA AND ACTION ITEMS:

### 1.1. Background:

- 1.1.1. The Sequence IV Surveillance Panel recently formed two sub-groups to address the open Sequence IVB action items.
  - 1.1.1.1. The first sub-group is tasked with improving test precision.
  - 1.1.1.2. The second sub-group is tasked with finalizing the test procedure.
- 1.1.2. Lubrizol recently recommended that a third sub-group be formed to work on the metrology-related action items.
  - 1.1.2.1. The members of the original two sub-groups agreed with this recommendation.

### 1.2. Current Metrology Action Items:

- 1.2.1. Develop a D.O.E. (Round-Robin) to compare the Keyence units at each of the five participating laboratories.
- 1.2.2. Provide input to the two original sub-groups regarding what metrology data should be included in the IVB test report.
- 1.2.3. Develop a procedure to use the Keyence G2 software to screen lifter profiles.
- 1.2.4. Determine whether the Keyence instruments should be monitored in LTMS.

### 1.3. Agenda for this Meeting:

- 1.3.1. Chris Mileti (Lubrizol) will lead this meeting.
  - 1.3.1.1. He will also compile minutes and action items for all future meetings.

#### 1.3.2. *The agenda for this meeting is simple:*

- 1.3.2.1. Agree on a frequency and format for future meetings.
- 1.3.2.2. Review the existing action items.
- 1.3.2.3. Prioritize the existing action items in terms of importance and difficulty to complete.
- 1.3.2.4. Determine if any "open" action items were accidentally omitted from the original list.
- 1.3.2.5. Begin assigning tasks to individual labs.

## 2. DISCUSSION:

### 2.1. Meeting Frequency and Format:

- 2.1.1. The attendees agreed to have a conference call every two weeks.
- 2.1.2. Lubrizol committed to issuing agendas, meeting minutes and action items.
- 2.1.3. Afton will be unable to participate in any conference calls between July 10<sup>th</sup> and July 12<sup>th</sup>.
  - 2.1.3.1. The Afton metrology technicians will be at Lubrizol's facility (Cleveland, OH) for the Gear Rating Workshop.

## **2.2. Additional Action Items:**

- 2.2.1. There was a consensus among the group that there are no additional action items to add to the list.

## **2.3. Develop D.O.E. to Compare All Five Keyence Units:**

### **2.3.1. Lubrizol Review of Previous Round-Robins:**

- 2.3.1.1. The three original labs (Intertek, Lubrizol and Southwest) previously conducted three separate round-robin experiments.
- 2.3.1.2. These experiments were designed by Kevin O'Malley (Statistics Group).
- 2.3.1.3. Statistical differences between labs were identified.
- 2.3.1.4. The results from these three round-robins should be reviewed by this sub-group.

### **2.3.2. Pre-Test and Post-Test Measurements:**

- 2.3.2.1. There was a lot of debate as to whether only pre-test measurements should be included in the round-robin.
- 2.3.2.2. Limiting the round-robin to pre-test measurements only would make the associated logistics much simpler.
- 2.3.2.3. However, using both pre-test and post-test measurements would provide a stronger tie-back to the test itself (because the actual pass/fail parameter is volume loss).
- 2.3.3. There was concern within the group as to how a Keyence unit will be "fixed" if it is found to be different than the other four macroscopes.
  - 2.3.3.1. The manufacturer must send the macroscopes back to Japan for repairs.

### **2.3.4. LTMS System:**

- 2.3.4.1. The statisticians have proposed adding the Keyence units to LTMS.
- 2.3.4.2. Periodic verification data could be uploaded to LTMS to monitor for severity and performance shifts.
- 2.3.4.3. One of the benefits to using LTMS is that small shifts in instrument performance could (theoretically) be handled by severity adjustments instead of repairs.

### **2.3.5. BOI/VGRA Matrix:**

- 2.3.5.1. Intertek may be willing to use one of their new BOI/VGRA engines to generate the post-test lifters for the round-robin.
- 2.3.5.2. Mileti, Coker and Buscher will discuss this possibility and then report back to the sub-group.

### **2.3.6. Forward Action Plan:**

- 2.3.6.1. Mileti would like to prioritize this action item because it will take the most time to complete.
  - 2.3.6.1.1. It will be the focus of the next sub-group meeting.
  - 2.3.6.1.2. Representatives from the Statistics Group will be invited to attend.

## **2.4. Use Keyence to Screen Lifters for Profile Quality:**

### **2.4.1. Explanation from Lubrizol:**

- 2.4.1.1. Toyota commented on this issue during a recent sub-group meeting.
  - 2.4.1.1.1. A lifter will rotate if its profile is crowned or flat.
  - 2.4.1.1.2. A lifter will not rotate if its profile is concave.

- 2.4.1.2. As a result, the Keyence will need to eliminate any lifter that has a concave profile.
- 2.4.1.3. The Surveillance Panel has changed its position several times within the last year regarding how strict the rejection criteria should be.

**2.4.2. High-Spots:**

- 2.4.2.1. The group discussed how high-spots, or unusual peaks, in pre-test lifter profiles should be handled.
- 2.4.2.2. Some high-spots are due to an anomaly that occurs during imaging.
  - 2.4.2.2.1. These high-spots can be eliminated by re-measuring the part.
- 2.4.2.3. Other high-spots are physical peaks in the surface of the lifter.
  - 2.4.2.3.1. Most labs will reject lifters that exhibit these irregularities.

**2.4.3. Forward Action Plan:**

- 2.4.3.1. Most of the labs already use their own method to screen lifters.
- 2.4.3.2. Each lab has been asked to document their current method in an email to the entire sub-group.
  - 2.4.3.2.1. Mileti will then summarize all this feedback so that it can be reviewed in a future sub-group meeting.
- 2.4.3.3. Each lab has also been asked to provide images that illustrate what they consider to be unacceptable lifter profiles.

**2.5. Calibrating/Verifying Keyence Units:**

- 2.5.1. There was a lot of discussion regarding how to verify the performance of a Keyence unit.
- 2.5.2. Keyence supplies a verification block with each macroscope.
  - 2.5.2.1. This block could be used, although it does not directly represent the measurements taken during a IVB test.
- 2.5.3. Another option would be for each lab to maintain a dedicated set of lifters that are used exclusively for taking verification measurements.
- 2.5.4. Mileti will devote an entire sub-group conference call specifically to this topic.

Action Items	Person responsible	Completion Date

Follow-up Notes/Updates	Initials	Date Added
Representatives from each of the (5) Sequence IVB laboratories participated in this conference call.	CHTM	07-02-2018

Attendees	Organization	Contact Information

Attendees	Organization	Contact Information