



Test Monitoring Center

6555 Penn Avenue
Pittsburgh, PA 15206-4489
(412) 365-1000

MEMORANDUM: 02-119
DATE: November 25, 2002
TO: L-33 Surveillance Panel
FROM: Donald Lind
SUBJECT: V01.1 Hardware Approval

At the November 20, 2002 L-33 Surveillance Panel meeting, the panel approved the new V01.1 hardware for use with the existing reference oil targets established on V99.1 hardware shown in the tables below.

Reference oil 123 Test Targets		
Parameter	Mean	Standard Deviation
RUST	8.43	0.390

Reference oil 151-3 Test Targets		
Parameter	Mean	Standard Deviation
RUST	9.69	0.350

These targets are effective for all tests completed on or after November 25, 2002.

DML/dml

c: Frank Farber, TMC
John L. Zalar, TMC
David DuBois, PRI
<ftp://ftp.astmtmc.cmu.edu/docs/gears/l33/memos/mem02-119.pdf>

Distribution: email