



# Test Monitoring Center

6555 Penn Avenue  
Pittsburgh, PA 15206-4489  
(412) 365-1000

MEMORANDUM: 02-120  
DATE: November 25, 2002  
TO: L-33 Surveillance Panel  
FROM: Donald Lind  
SUBJECT: Reference Oil 123-2 Approval

At the November 20, 2002 L-33 Surveillance Panel meeting, the panel approved the use of reference oil 123-2 with V99.1 hardware. The targets, shown in the table below, were established from the existing reference oil 123 targets conducted on V99.1 hardware.

<b>Reference oil 123-2 Test Targets</b>		
<b>Parameter</b>	<b>Mean</b>	<b>Standard Deviation</b>
RUST	8.43	0.390

These targets are effective for all tests completed on or after November 25, 2002.

DML/dml

c: Frank Farber, TMC  
John L. Zalar, TMC  
<ftp://ftp.astmtmc.cmu.edu/docs/gears/l33/memos/mem02-120.pdf>

Distribution: email