



Test Monitoring Center

@ Carnegie Mellon University
6555 Penn Avenue, Pittsburgh, PA 15206, USA

<http://astmtmc.cmu.edu>
412-365-1000

MEMORANDUM: 20-044

DATE: October 12, 2020

TO: Matthew Sangpeal, Chairman, L-42 Surveillance Panel

FROM: Dylan Beck *Dylan Beck*

SUBJECT: L-42 Rater Calibration from April 1, 2020 through September 30, 2020

Attached is a summary of L-42 rater calibration activity this period.

DJB/djb/mem20-044.djb.doc


cc: Frank Farber
Jeff Clark
L-42 Surveillance Panel

<http://www.astmtmc.cmu.edu/ftp/docs/gear/l42rc/semiannualreports/l42rc-10-2020.pdf>

Distribution: email

L-42 Rater Calibration

	Reporting Data	Calibrated on 9-30-20
Number of Labs	4	4
Number of Raters	12	12



11:45:24 09OCT2020

L-42 Rater Calibration

Test Distribution by Validity

		Totals	
		Last Period	This Period
Accepted for calibration	AC	17	15
Rejected (Mild)	OC	0	0
Rejected (Severe)	OC	1	0
Rejected (Multiple)	OC	0	0
Rejected (Precision)	OC	0	0
Workshop data	AG	10	0
Total		28	15


L-42 Rater Calibration

Rater	N	PSCR		RSCR	
		Avg Yi	STD*	Avg Yi	STD*
AA	8	-0.499	1.266	-0.151	1.077
BB	8	0.251	0.821	0.279	0.786
C	8	0.487	0.768	-0.049	0.897
H	4	0.068	0.643	-0.086	0.815
I	4	0.043	0.310	-0.446	0.761
N	4	0.160	0.267	0.077	0.439
O	4	-0.036	0.777	0.716	0.895
P	4	-0.119	0.159	0.151	0.438
Q	4	-0.188	0.346	-0.150	0.314
U	4	0.104	0.704	-0.584	0.254
V	4	-0.014	0.069	0.031	0.151
Z	4	-0.837	0.484	-1.185	0.766
ALL	60	-0.023	0.757	-0.088	0.804

* Due to the small number of ratings per part, the standard deviation of the Yi values is given in place of a pooled standard deviation.

L-42 Rater Calibration


PINION SCORING
Severity as Measured by Delta/s



11:45:24 09OCT2020

L-42 Rater Calibration

RING SCORING
Severity as Measured by Delta/s




11:45:24 09OCT2020

L-42 Rater Calibration

PINION SCORING

Precision as Measured by \bar{Y}_i STD




11:45:24 09OCT2020

L-42 Rater Calibration

RING SCORING

Precision as Measured by Yi STD



11:45:24 09OCT2020

L-42 Rater Calibration

SUMMARY OF SEVERITY & PRECISION

Severity

PSCR remained within limits this period. RSCR exceeded the action limit with one test this period but finished the period within limits.

Precision


Precision for both parameters remained within acceptable limits.

Industry control charts follow.

L-42 Rater Calibration

L-42 RATER CALIBRATION INDUSTRY OPERATIONALLY VALID DATA


PINION SCORING



L-42 Rater Calibration

L-42 RATER CALIBRATION INDUSTRY OPERATIONALLY VALID DATA

RING SCORING




L-42 Rater Calibration

L-42 RATER CALIBRATION INDUSTRY OPERATIONALLY VALID DATA

PINION SCORING

LTMS Precision Analysis




09OCT20:13:19

L-42 Rater Calibration

L-42 RATER CALIBRATION INDUSTRY OPERATIONALLY VALID DATA

RING SCORING

LTMS Precision Analysis



L-42 Rater Calibration

L-42 RATER CALIBRATION INDUSTRY OPERATIONALLY VALID DATA

PINION SCORING

CUSUM Severity Analysis



L-42 Rater Calibration

L-42 RATER CALIBRATION INDUSTRY OPERATIONALLY VALID DATA

RING SCORING

CUSUM Severity Analysis



09OCT20:13:20