

**Report of Meeting**  
**L-60-1 Surveillance Panel Conference Call**  
**August 10th, 2022**

**Attendees:**

SwRI -	<b>Mueller</b> , Kostan, Charron
Lubrizol -	Venhoff, <b>Slocum</b> , Schaup
Afton -	<b>Sangpeal</b> , Bell, Horvath
Intertek -	<b>Lange</b>
TMC -	<b>Beck</b>
ExxonMobil -	<b>Banas</b>
BASF -	<b>Goyal</b> , Mosher, Caridi
Dana -	<b>Zyski</b>
Meritor -	LaBond, <b>Carter</b>
Army -	Sattler
AAM -	<b>Muransky</b>
Shell -	Uy
Chevron -	Warden
Daimler -	Neal

Voting Members in **BOLD**

**1.0 Membership Review**

- Remove Mike Cabaj

**2.0 Meeting minutes Approval**

- May 11th, 2022, ASTM Meeting #205

**Motion #1 → R. Slocum 1<sup>st</sup> /2<sup>nd</sup> M. Sangpeal to approve the meeting minutes from the May 11th, 2022, ASTM Meeting. Motion passed unanimously, 10-0-0 (Yes-No-Abstain).**

**3.0 L-60-1 Severity Task Force Follow-up**

- Labs will run same bar stock, same reference oil (148), and gears prepped by Afton
- Upon shipment to labs gears will be coated in reference oil, wrapped in Nox Rust paper, and put in plastic bag
- TMC will send labs fresh reference oil samples
- Aim to run mid-September
- LZ/Intertek still need to generate PO's to Prosimalys for gear analysis possibly complete end of September
  - When complete follow-up presentation from Prosimalys and the severity task force
- Intertek to upload reference data sets
  - Todd Dvorak was going to look at data but no longer available but will involve Travis

#### 4.0 Sec. 6.1.8.2 venturi flow meter calibration

**Motion #2** → A. Lange 1<sup>st</sup> /2<sup>nd</sup> W. Venhoff to approve the change to Sec. 6.1.8.2 from D5704-20 to the following –

**6.1.8.2** Send the Preso Low Loss Venturi Meter together with the Dwyer digital manometer to the specified calibration laboratory<sup>16</sup> **or any ISO/IEC 17025 accredited laboratory** for cleaning and calibration at least once a year.

**Motion passed unanimously, 10-0-0 (Yes-No-Abstain).**

#### Action Items:

- R. Slocum to investigate into addressing or removing “Sole Source Suppliers” instances in D5704

#### 5.0 Adjourn

**Motion #3** → T. Muransky 1<sup>st</sup> /2<sup>nd</sup> A. Lange to adjourn. **Motion passed unanimously, 10-0-0 (Yes-No-Abstain).**

Respectfully submitted,

Robert Slocum  
L-60-1 Surveillance Panel Chairman



D02.B0.03

# L-60-1 Surveillance Panel Meeting

8/10/2022

14:00 pm– 15:00 pm

Robert Slocum

# Agenda

- Call to Order/Agenda review
- Membership review
- Meeting Minute Approvals
  - May 11<sup>th</sup>, 2022, ASTM Meeting
- L-60-1 Severity Task Force Follow-up
  - 4 lab run
  - Gear analysis
  - TMC data upload
- Sec. 6.1.8.2 venturi flow meter calibration
  - Other Calibration Labs?
- Old Business
- New business
- Adjournment

# Membership Review

<b>Allen Comfort</b>	<b>US Army</b>
<b>Amy Zyski</b>	<b>Dana</b>
<b>Arjun Goyal</b>	<b>BASF</b>
<b>Anthony Lange</b>	<b>Intertek</b>
<b>Jason W. Carter</b>	<b>Meritor</b>
<b>Dylan Beck</b>	<b>TMC</b>
<b>Robert Slocum</b>	<b>Lubrizol</b>
<b>Matt Sangpeal</b>	<b>Afton</b>
<b>Mike Cabaj</b>	<b>Linamar</b>
<b>Caroline Mueller</b>	<b>SwRI</b>
<b>Rob Banas</b>	<b>ExxonMobil</b>
<b>Troy Muransky</b>	<b>AAM</b>

# Meeting Minutes Approval

– May 11<sup>th</sup>, 2022, ASTM Meeting

# L-60-1 Severity Task Force Follow-up

- 4 lab run
  - Stands available and timing?
- Metal analysis
  - PO's??
- TMC lab reference data upload







## – Borderline Oil Data

Stand	Run #	Oil	Small Gear Batch	Large Gear Batch	AVG C/V w/CF	Lg Gr Avg	Lg Gr F	Lg Gr R	Sm Gr F
4F	378	GO-013357-07-00	12-11-11	12-11-45	6.80	6.20	6.00	6.40	6.50
4F	386	GO-013357-07-02	07-18-30	06-18-47	5.40	4.80	4.70	5.00	5.61
5F	435	GO-013357-07-01	12-11-05	12-11-36	6.40	5.80	5.40	6.11	3.34
5F	440	GO-013357-07-02	07-18-30	07-18-47	5.60	5.00	4.75	5.30	0.98
16	228	GO-013357-07-01	06-18-22	05-18-60	7.60	7.00	6.90	7.05	6.95
16	233	GO-013357-07-02	06-18-22	05-18-65	7.30	6.70	6.85	6.50	6.40
16	239	GO-013357-07-02	12-11-07	12-11-45	*7.1	6.50	6.80	6.25	6.41
					*S.A. Removed				

# L-60-1 Severity Initial Summary

## Executive Summary

### Average Carbon/Varnish

- Reference oil data historically has been consistently severe of target.
- Though initial testing on 2018 hardware was closer to target, all labs who have switched have seen a shift, and the one lab that hasn't has remained stable, therefore indicating a high likelihood of the new gear batch as the cause of the shift.
- Both reference oils have shifted similarly, providing evidence that candidate behavior is likely also shifted.
- Options:
  1. Correction factor 0.6 merits is recommended for a tests run with the 2018 hardware, in addition to the 0.6 merit correction currently in place, for a total correction of 1.2 merits for tests run on the 2018 hardware.
  2. There are clear differences in precision of ACV across labs, with Lab D running much more precise than other labs. Lab visits are recommended to better understand these differences.
    - Updates to the ACV correction factor may be delayed - pending the findings / conclusions of the lab visits
  3. Do nothing

### Average Sludge

- Only 3 data points beyond 2.0 sigma severe, and there is no clear shift at the time of the new hardware introduction.
- It is recommended to continue to monitor this parameter without a correction factor and allow severity adjustments to handle differences in severity.

- Sec. 6.1.8.2 venturi flow meter calibration

6.1.8.2 Send the Preso Low Loss Venturi Meter together with the Dwyer digital manometer to the specified calibration laboratory<sup>16</sup> for cleaning and calibration at least once a year.

(16) Bowser-Morner, 4518 Taylorsville Rd.,  
Dayton, OH 45424.

6.1.8.2 Send the Preso Low Loss Venturi Meter together with the Dwyer digital manometer to the specified calibration laboratory<sup>16</sup> or any ISO/IEC 17025 accredited laboratory for cleaning and calibration at least once a year.

Old Business

New Business

Adjournment