



Test Monitoring Center

6555 Penn Avenue
Pittsburgh, PA 15206-4489
(412) 365-1000

L-60-1 Information Letter 06-1
Sequence Number 32
July 11, 2006

ASTM consensus has not been obtained on this information letter. An appropriate ASTM ballot will be issued in order to achieve such consensus.

TO: L-60-1 Mailing List

SUBJECT: 1. Revised Copper Catalyst Strip Cleaning Procedure
2. Editorial Revision

1. At the June 21, 2006 L-60-1 Surveillance Panel meeting, the panel approved a motion to revise the copper catalyst strip cleaning procedure. This revision changes the 30 minute time restriction to approximately 30 minutes. A revised Section 12.1.1 of Test Method D5704 is attached. This change is effective the date of this information letter.

2. At the June 21, 2006 L-60-1 Surveillance Panel meeting, the panel approved a motion to delete Section 6.1.8.3 since it repeats the actions in Section 6.1.8.1. This change is effective the date of this information letter.

Chris Schenkenberger
Chairman
L-60-1 Surveillance Panel

John L. Zalar
Administrator
ASTM Test Monitoring Center

Attachment

c: ftp://ftp.astmtmc.cmu.edu/docs/gears/l601/procedure_and_ils/il06-1.pdf

Distribution: Email

(Revises Test Method D 5704-06)

12.1.1 Carefully remove all the deposits from the notched copper catalyst strip by soaking for approximately 30 min in Oakite 811, Penmul L460, or equivalent.

DELETE Section 6.1.8.3