



Test Monitoring Center

203 Armstrong Drive, Freeport, PA 16229, USA

www.astmtmc.org
412-365-1000

L-60-1 Information Letter 22-3
Sequence Number 53
August 29, 2022

ASTM consensus has not been obtained on this information letter. An appropriate ASTM ballot will be issued in order to achieve such consensus.

TO: L-60-1 Surveillance Panel

SUBJECT: Change to Sole Sourced Calibration Laboratory

On August 10, 2022, the L-60-1 Surveillance Panel approved modifications to the test procedure to eliminate a single laboratory as the specified provider to clean and calibrate instruments used to conduct testing. Previously Bowser-Morner in Dayton, OH was the only specified laboratory. This became an issue when Bowser-Morner had longer than expected lead times to provide the required services. The change to the test procedure will approve any ISO/IEC 17025 accredited laboratory to provide these services. This change is effective immediately following the conclusion of the surveillance panel vote.

Robert Slocum
Chairman
L-60-1 Surveillance Panel

Frank M. Farber
Director
ASTM Test Monitoring Center

Attachments

cc: https://www.astmtmc.org/docs/gear/l601/procedure_and_ils/il22-3_L601.pdf

Distribution: Email

(Revises Test Method D 5704-22)

Replace the text of section 6.1.8.2 and footnote 16 with the following:

6.1.8.2 Send the Preso Low Loss Venturi Meter together with the Dwyer digital manometer to the specified calibration laboratory¹⁶ for cleaning and calibration at least once a year.

¹⁶ ~~Bowser Morner, 4518 Taylorsville Rd., Dayton, OH 45424.~~ Any ISO/IEC 17025 accredited laboratory.