

Oil Seal Compatibility Test (OSCT) Surveillance Panel Meeting Minutes

06/16/2009

D. Bell

Meeting Attendance:

D. Bell (Afton)
J. Gropp (Lz)
D. Lind (TMC)
J. Mattern (Lz)
C. Knight (TEI)
J. Keiter (Lz)
B. McGlone (ArvinMeritor)
B. Grinfield (SWRI)
M. Kasimirsky (TMC, not present, but voted prior to the meeting)

The qualification data compiled by the Test Monitoring Center that was generated by Lz and SWRI for Polyacrylate batch PA340 in reference oils 160-1 and 161-1 were reviewed by the OSCT Surveillance Panel. The shore hardness, percent elongation, and % volume change for PA340 were well within the acceptance bands in both reference oils and in good agreement between labs. A motion made by B. Grinfield and 2nd by B. McGlone passed with 6 approvals, 0 disapprovals, and 0 abstentions to approve PA340 effective 6/16/09 for OSCT per ASTM D5662. The elastomer distributor, Test Engineering Institute (TEI), was notified to release PA340 (~200 slabs) for sale to SWRI and Lz as orders are received.

Motion to adjourn was unanimously approved on 6/16/09 at 10:12 am.